



APPLICATION NOTE NO. 11 General

June 2005

PAR Light Sensors

Sea-Bird has several application notes dealing with PAR sensors from various manufacturers; this application note provides an overview of PAR measurements and units, and is applicable to all PAR sensors.

PAR is an abbreviation for **Photosynthetically Available Radiation** (also called **Photosynthetically Active Radiation**). Solar radiation reaching the Earth's surface comprises a mixture of ultraviolet light, visible light, and near-visible infrared radiation. All of this radiation conveys heat; the portion between approximately 400 nm and 700 nm wavelength can be captured and used by photo-autotrophs (organisms that are capable of obtaining their energy directly from sunlight), and is called PAR.

Irradiance is the flux of solar radiation incident on a surface per unit time per unit area and is reported in units of energy content (Watts/m²) or photon content (quanta/m² sec or μEinsteins/m² sec). Conversion from energy to photon content can be made with Planck's equation, provided that the light wavelength is known. The energy of a photon is related to its wavelength as follows:

E = hc / λ
where
 h = Planck's constant (6.626 x 10⁻³⁴ Joules/sec)
 c = speed of light (2.998 x 10⁸ m/sec)
 λ = wavelength (m)

This equation provides the energy for a single wavelength. For a broad spectrum PAR sensor, a wavelength of approximately 550 nm (550 x 10⁻⁹ m) is typically used for the conversion.

"For marine atmospheres with sun altitudes above 22 degrees, the quanta/watt ratio for the region 400 to 700 nm is 2.77 x 10¹⁸ quanta/sec/Watt to an accuracy of plus or minus a few percent." This quote and further discussion of the relationship of quanta to Watts in the water column is found in Smith and Morel (1974) Limnol. Oceanogr. 19(4):591-600.

E (at 550 nm) = hc / λ = (6.626 x 10⁻³⁴ Joules/sec) * (2.998 x 10⁸ m/sec) / (550 x 10⁻⁹ m) = 3.61 x 10⁻¹⁹ Joules
 (Note: 1 / 3.61 x 10⁻¹⁹ = 2.77 x 10¹⁸ quanta/sec/Watt, the value quoted in the above reference.)

Application notes for underwater PAR sensors (**11Chelsea, 11Licor, 11QSP-L, and 11QSP-PD**) and surface PAR sensors (**11S and 47**) describe how to enter coefficients from the manufacturer's calibration in the CTD configuration (.con) file to provide SEASOFT output in μEinsteins/m²·sec. To calculate irradiance in other units:

To convert to:	For Underwater PAR Sensors, set Multiplier to:	For Surface PAR Sensors, multiply calculated Conversion factor by:
μEinsteins/m ² ·sec	1.0	
μEinsteins/cm ² ·sec	(1.0) / (100 cm/m) ² = 1 x 10⁻⁴	
Einsteins/m ² ·sec	(1.0) / (1 x 10 ⁶ μEinsteins/Einstein) = 1 x 10⁻⁶	
Einsteins/cm ² ·sec	(1 x 10 ⁻⁶) / (100 cm/m) ² = 1 x 10⁻¹⁰	
quanta/m ² ·sec	(1 x 10 ⁻⁶) * (6.022 x 10 ²³ quanta/Einstein) = 6.022 x 10¹⁷	
quanta/cm ² ·sec	(6.022 x 10 ¹⁷) / (100 cm/m) ² = 6.022 x 10¹³	
Watts/m ²	(6.022 x 10 ¹⁷) / (2.77 x 10 ¹⁸ quanta/sec/Watt) = 0.2174	
Watts/cm ²	(0.2174) / (100 cm/m) ² = 2.174 x 10⁻⁵	
μWatts/m ²	(0.2174) * (1 x 10 ⁶ μWatts/Watt) = 2.174 x 10⁵	

Note: 1 Einstein = 1 mole (6.022 x 10²³) of photons 1 Watt = 2.77 x 10¹⁸ quanta/sec

Notes:

- Edit the CTD configuration (.con) file using the Configure menu (in SEASAVE or SBE Data Processing in our SEASOFT-Win32 suite of programs) or SEACON (in SEASOFT-DOS).
- Multiplier can also be used to scale output for comparing the shape of data sets taken at disparate light levels. For example, a multiplier of 10 would make a 10 μEinsteins/m²·sec light level plot as 100 μEinsteins/m²·sec.