

SUNA1918

Version: V2

Operator:

Reference Update Time: 07-Sep-2022 23:05:53

Comment: Summary Report created automatically by the UCI Diagnostic Data Collection tool.

Ancillary	Value	Setting	Value
Firmware Revision	2.5.1	SerialBaudRate	57600
Total Disk Space (byte)	3954802688	OperationMode	Periodic
Free Disk Space (byte)	3760291840	FixedDuration	60
Used Lamp Time (sec)	211885	FrameTransfer	NONE
Clock Time	14 Jul 2023 01:14:28 UTC	FrameLogging	FULL_ASCII
USB Enabled	true	OperationControl	Samples
PCB Supervisor	true	CycleLightFrames	29
Analog Board	true	CycleDarkFrames	1
SDI12 Board	true	CycleLightDuration	58
Relay Board	false	CycleDarkDuration	2
Super Capacitors	true	Baseline	Linear
Scheduling	true	Adaptive Integration	On
Data Logging	true	Concentrations Fitted	N03_SW_TSW
APF Interface	false	Concentrations Available	3
Integrated Wiper	true	Bromide Trace	false
External Power Port	false	IntegPeriod (ms)	225
Lamp Switch Off Temperature	35	WavelenMin (nm)	217.0
		WavelenMax (nm)	240.0
		NO3DacMin (μM)	-5.0
		NO3DacMax (μM)	100.0
		SpectralInLightAvg	1
		SpectralInDarkAvg	1
		PeriodicInterval	1 hr
		PeriodicOffset	0
		PeriodicDuration	5
		PeriodicSamples	15
		LogFileType	Daily
		LogFileMaxSz	5

SUNA 1918

Version: V2

Configuration

SENSTYPE SUNA
SENSVERS V2
SERIALNO 1918
THEBRAND SeabirdCoastal
PATHLGTH 10mm
INTWIPER Available
EXTPPORT Missing
SUPRCAPS Available
PWRSVISR Available
USBSWTCH Available
RELAYBRD Missing
SDI12BRD Available
ANALGBRD Available
INTDATLG Available
APFIFACE Missing
SCHDLING Available
OWIRETLP 10eb08b703080091
OWIRETSP 108825b103080016
OWIRETHS 10d13ec10308000a
ZSPEC_SN 141177
FIBERLSN L21.E01.212664
CUSTOMID None
STUPSTUS Done
BAUDRATE 57600
MSGLEVEL Info
MSGFSIZE 2
DATFSIZE 5
OUTFRTYP None
LOGFRTYP Full_ASCII
OUTDRKFR Output
LOGDRKFR Output
LOGFTYPE Daily
AFILEDUR 60
ACQCOUNT 5
CNTCOUNT 0
DCMINNO3 -5.000
DCMAXNO3 100.000
WDAT_LOW 217.00
WDAT_HGH 250.00
SDI12ADD 48
DATAMODE Real
OPERMODE Periodic
OPERCTRL Samples
EXDEV TYP Wiper
EXDEVPRE 15
EXDEV RUN Off
EXDVIVAL 0

Configuration

COUNTDWN 3
FIXDDURA 60
PERDIVAL 1h
PERDOFFS 0
PERDDURA 5
PERDSMPL 15
POLLTOUT 15
APFATOFF 10.0000
STBLTIME 5
SKPSLEEP Off
LAMPTOFF 35
SPINTPER 225
DRKAVERS 1
LGTAVERS 1
DRKSMPLS 1
LGTSMPLS 29
DRKDURAT 2
LGTDURAT 58
TEMPCOMP Off
SALINFIT On
BRMTRACE Off
BL_ORDER 1
FITCONCS 3
DRKCORMT SpecAverage
A_CUTOFF 1.3000
INTPRADJ On
INTPRFAC 1
INTADSTP 5
INTADMAX 20
WFIT_LOW 217.00
WFIT_HGH 240.00
LAMPTIME 211885